

Liofilchem S.r.l.	CERTIFICATO CONTROLLO QUALITÀ	N° 78618 - 78619
	<i>QUALITY CONTROL CERTIFICATE</i>	Revisione/ Revision 7 _ 10.01.2014 Pag. 1 di 1

PRODOTTO / *PRODUCT* : **EnteroPluri-Test**

LOTTO / *BATCH*: **021119004**

DATA PRODUZIONE/*PRODUCTION DATE*: **11.02.2019**

DATA SCADENZA / *EXP. DATE* : **2019.10.09**

CARATTERISTICHE BIOLOGICHE / *BIOLOGICAL CHARACTERISTICS*:

Tempo di incubazione / <i>Incubation time</i>	18-24 ore/ <i>hours</i>
Temperatura / <i>Temperature</i>	36 °C ± 1 °C
Modalità di incubazione / <i>Procedure of incubation</i>	√ O ₂ CO ₂
Aggiunta di KOH 40% / <i>Adding of KOH 40%</i>	2 gocce/drops
Aggiunta di α-naftolo / <i>Adding of α-naphthol</i>	3 gocce/drops
Aggiunta di Kovac's / <i>Adding of Kovac's</i>	3-4 gocce/drops

	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15
<i>Microrganisms</i>	GLU	GAS	LYS	ORN	H2S	IND	ADO	LAC	ARA	SOR	VP	DUL	PA	UR	CIT
<i>Escherichia coli ATCC 25922</i>															
<i>Theoretical results</i>	+	±	+	+	-	+	-	+	+	+	-	-	-	-	-
<i>Results of conformity</i>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>
<i>Citrobacter freundii ATCC 8090</i>															
<i>Theoretical results</i>	+	+	-	±	+	-	-	±	+	+	-	±	-	-	±
<i>Results of conformity</i>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>
<i>Proteus mirabilis ATCC 25933</i>															
<i>Theoretical results</i>	+	±	-	+	+	-	-	-	-	-	±	-	+	+	±
<i>Results of conformity</i>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>
<i>Enterobacter cloacae ATCC 13047</i>															
<i>Theoretical results</i>	+	+	-	+	-	-	±	±	+	+	+	±	-	-	+
<i>Results of conformity</i>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>
<i>Klebsiella pneumoniae ATCC 13883</i>															
<i>Theoretical results</i>	+	±	+	-	-	-	+	+	+	+	±	+	-	±	+
<i>Results of conformity</i>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>
<i>Salmonella typhimurium ATCC 14028</i>															
<i>Theoretical results</i>	+	±	+	±	+	-	-	-	+	+	-	-	-	-	±
<i>Results of conformity</i>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>

+ Reazione positiva / *Positive reaction* **-** Reazione negativa / *Negative reaction* **±** Reazione variabile / *Variable reaction*

LOTTO / *BATCH* : Idoneo / *Approved*

Non Idoneo / *Not approved*

DATA / *DATE* : 14.02.2019

Responsabile Controllo Qualità
Quality Control Manager

(D. Vitagliano)

Dario Vitagliano