

<h1 style="margin: 0;">Liofilchem</h1> <h2 style="margin: 0;">S.r.l.</h2>	<b>CERTIFICATO CONTROLLO QUALITÀ</b>  <b>QUALITY CONTROL CERTIFICATE</b>	<b>N° 78618 - 78619</b>  Revisione/ Revision 7 _ 10.01.2014 Pag. 1 di 1

PRODOTTO / *PRODUCT* : **EnteroPluri-Test**

LOTTO / *BATCH*: **100818002**

DATA PRODUZIONE/*PRODUCTION DATE*: **08.10.2018**

DATA SCADENZA / *EXP. DATE* : **2019.08.04**

**CARATTERISTICHE BIOLOGICHE / *BIOLOGICAL CHARACTERISTICS*:**

Tempo di incubazione / <i>Incubation time</i>	18-24 ore/ <i>hours</i>
Temperatura / <i>Temperature</i>	36 °C ± 1 °C
Modalità di incubazione / <i>Procedure of incubation</i>	√ O <sub>2</sub> CO <sub>2</sub>
Aggiunta di KOH 40% / <i>Adding of KOH 40%</i>	2 gocce/drops
Aggiunta di α-naftolo / <i>Adding of α-naphthol</i>	3 gocce/drops
Aggiunta di Kovac's / <i>Adding of Kovac's</i>	3-4 gocce/drops

	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15
<i>Microrganisms</i>	GLU	GAS	LYS	ORN	H2S	IND	ADO	LAC	ARA	SOR	VP	DUL	PA	UR	CIT
<b><i>Escherichia coli ATCC 25922</i></b>															
<i>Theoretical results</i>	+	±	+	+	-	+	-	+	+	+	-	-	-	-	-
<i>Results of conformity</i>	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒
<b><i>Citrobacter freundii ATCC 8090</i></b>															
<i>Theoretical results</i>	+	+	-	±	+	-	-	±	+	+	-	±	-	-	±
<i>Results of conformity</i>	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒
<b><i>Proteus mirabilis ATCC 25933</i></b>															
<i>Theoretical results</i>	+	±	-	+	+	-	-	-	-	-	±	-	+	+	±
<i>Results of conformity</i>	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒
<b><i>Enterobacter cloacae ATCC 13047</i></b>															
<i>Theoretical results</i>	+	+	-	+	-	-	±	±	+	+	+	±	-	-	+
<i>Results of conformity</i>	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒
<b><i>Klebsiella pneumoniae ATCC 13883</i></b>															
<i>Theoretical results</i>	+	±	+	-	-	-	+	+	+	+	±	+	-	±	+
<i>Results of conformity</i>	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒
<b><i>Salmonella typhimurium ATCC 14028</i></b>															
<i>Theoretical results</i>	+	±	+	±	+	-	-	-	+	+	-	-	-	-	±
<i>Results of conformity</i>	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒

+ **Reazione positiva / *Positive reaction***      - **Reazione negativa / *Negative reaction***      ± **Reazione variabile / *Variable reaction***

LOTTO / *BATCH* : ☒ Idoneo / *Approved*

☐ Non Idoneo / *Not approved*

DATA / *DATE* : 11.10.2018

Responsabile Controllo Qualità  
*Quality Control Manager*

(D. Vitagliano)

*Dario Vitagliano*