

PRODOTTO / PRODUCT : **Enteropluri-Test**

LOTTO / BATCH: **101018037**

DATA PRODUZIONE/PRODUCTION DATE: **10.10.2018**

DATA SCADENZA / EXP. DATE : **2019.08.06**

CARATTERISTICHE BIOLOGICHE / BIOLOGICAL CHARACTERISTICS:

Tempo di incubazione / Incubation time	18-24 ore/ hours													
Temperatura / Temperature	36 °C ± 1 °C													
Modalità di incubazione / Procedure of incubation	\checkmark O ₂ CO ₂													
Aggiunta di KOH 40% / Adding of KOH 40%	2 gocce/drops													
Aggiunta di α -naftolo / Adding of α -naphthol	3 gocce/drops													
Aggiunta di Kovac's / Adding of Kovac's	3-4 gocce/drops													

	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15
Microrganismi	GLU	GAS	LYS	ORN	H2S	IND	ADO	LAC	ARA	SOR	VP	DUL	PA	UR	CIT
Escherichia coli ATCC 25922															
Theoretical results	+	±	+	+	-	+	-	+	+	+	-	-	-	-	-
Results of conformity	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒
Citrobacter freundii ATCC 8090															
Theoretical results	+	+	-	±	+	-	-	±	+	+	-	±	-	-	±
Results of conformity	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒
Proteus mirabilis ATCC 25933															
Theoretical results	+	±	-	+	+	-	-	-	-	-	±	-	+	+	±
Results of conformity	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒
Enterobacter cloacae ATCC 13047															
Theoretical results	+	+	-	+	-	-	±	±	+	+	+	±	-	-	+
Results of conformity	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒
Klebsiella pneumoniae ATCC 13883															
Theoretical results	+	±	+	-	-	-	+	+	+	+	±	+	-	±	+
Results of conformity	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒
Salmonella typhimurium ATCC 14028															
Theoretical results	+	±	+	±	+	-	-	-	+	+	-	-	-	-	±
Results of conformity	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒

+ Reazione positiva / Positive reaction

- Reazione negativa / Negative reaction

± Reazione variabile / Variable reaction

LOTTO / BATCH : Idoneo / Approved

Non Idoneo / Not approved

DATA / DATE : 15.10.2018

Responsabile Controllo Qualità
Quality Control Manager
(D. Vitagliano)

Dario Vitagliano