

Biolab Inc.
1141 Budapest, Öv u.43.
HUNGARY
Phone:+36-1 221-96-14
Fax +36-1 364 20 06
www.biolab.hu

Certificate of analysis

Name:	KOVÁCS INDOLE REAGENT for diagnostics purpose
Code Number:	KIR60100, KIR030615126
Lot:	KIR250917162
Manufacture Date:	09/2017
Expiry Date:	09/2018

Prepared Reagent:

Clarity:	Transparent
Deposit:	None

Microbiological Assay:	According to instruction for use
-------------------------------	----------------------------------

Quality control

Test strains	Indole reaction
<i>Escherichia coli</i>	Positive
<i>Salmonella enteritidis</i>	Negative

This product has been tested by Quality Control Laboratory and conforms to the specification contained in the relevant catalogue or to the specification agreed with the customer.

This product was manufactured by Biolab Inc.

Hereby we declare:

Under our sole responsibility, that the above mentioned devices meet the applicable provisions of the Directive 98/79/EC on „In Vitro Diagnostic Medical Devices”.

All the supporting documents, as required by Annex III on the 98/79/EC Directive, in order to prove conformity to the Essential Requirements as listed in Annex I, are retained under the premises of the Manufacturer.

This certificate was issued electronically and is valid without signature.