

Certificate of Quality


Product Name: RapID SS/u System KT/20
Lot Number: 349756

Product Number: R8311004
Expiration Date: 2019-05-29
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Inc., a part of Thermo Fisher Scientific Quality Control specifications and were found to meet performance criteria for this product.

Organism Challenge	ATCC	Performance Criteria
Escherichia coli	25922	Acceptable
Proteus mirabilis	29906	Acceptable
Enterococcus faecalis	29212	Acceptable
Klebsiella pneumoniae	13883	Acceptable
Serratia marcescens	8100	Acceptable

Signed



Quality Control Manager