

Certificate of Quality


Product Name: RapID SS/u System KT/20
Lot Number: 353501

Product Number: R8311004
Expiration Date: 2019-07-18
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Inc., a part of Thermo Fisher Scientific Quality Control specifications and were found to meet performance criteria for this product.

| Organism Challenge | ATCC | Performance Criteria |
|---------------------------|-------------|-----------------------------|
| Escherichia coli | 25922 | Acceptable |
| Proteus mirabilis | 29906 | Acceptable |
| Enterococcus faecalis | 29212 | Acceptable |
| Klebsiella pneumoniae | 13883 | Acceptable |
| Serratia marcescens | 8100 | Acceptable |

Signed



Quality Control Manager