

CERTIFICATE OF ANALYSIS

PRODUCT PO5090A CHOCOLATE AGAR + VITOX
LOT NUMBER 2424591
EXPIRY DATE 2019.01.28

General Characteristics	Results	Specification
Colour	Conforms	Fawn brown
Appearance	Conforms	Opaque
pH	7.0	7.0 - 7.4
Packaging / Presentation	Conforms	Conforms to specification
Sterility @ 25 & 36 ± 1°C for 72 hours	Conforms	Within Limits

Microbiological Performance	Control c.f.u	Test Result	Specification
Haemophilus influenzae Typ b ATCC®10211	62	56	3-5mm, cream shiny colonies
Neisseria gonorrhoeae ATCC®49226	<10000	Conforms	Good growth, cream shiny colonies
Staphylococcus epidermidis ATCC®12228	<10000	Conforms	Good growth, white colonies

For positive strains, colony count is greater than or equal to 70% of the control medium.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were obtained at the time of release.

This certificate is produced electronically and valid without a signature

The quality control methods meet requirements of ISO 11133:2014.