



Certificate of Analysis

Product: RAPID-SCAN 6-SN + **Art.-No.:** 01-04-11
Filling: 100 strips/bottle

Lot number: 0508/8376
Manufacturing date: 2019-01-25
Expiry date: 2021-01
Storage: Store between +2 and +30 °C.

Item	Specifications	Results	QC
Functional test			
<i>Glu - Pro - Bld - Nit - pH - SG</i>			
Negative/normal test	negative/normal	negative/normal	passed
Test with different concentrations/contents	Colouration corresponds to label, after the specified reading time (60 / 60-120 s)	corresponds	passed
Strip test			
Strip	Test-pads are well fixed on foil	well fixed	passed
	Foil is clean	clean	passed
	Right sequence of test-pads	correct	passed
Material test			
Desiccant cap	Desiccant exists	exists	passed
	Cap closes tight	tight	passed
Tube and Label	No damages and contamination	intact	passed
	Correct and well legible print	correct	passed
Sleeve	Correct position and shrinkage	correct	passed

This lot complies with our specification, in accordance with GMP.

Date of release: 2019-01-28 / Production Site Lichtenfels, Germany

This certificate of analysis has been automatically generated and is valid without signature.

Elena Schwietz
Manager QC Urine Diagnostics