

Certificate of Analysis

Product: RAPID-SCAN 8SL + **Art.-No.:** 01-04-21
Filling: 100 strips/bottle

Lot number: 0338/8000
Manufacturing date: 2018-08-24
Expiry date: 2020-08
Storage: Store between +2 and +30 °C.

| Item | Specifications | Results | QC |
|---|--|-----------------|--------|
| Functional test | | | |
| <i>Asc - Glu - Pro - Bld - pH - Nit - Leu -SG</i> | | | |
| Negative/normal test | negative/normal | negative/normal | passed |
| Test with different concentrations/contents | Colouration corresponds to label, after the specified reading time (60 / 60-120 s) | corresponds | passed |
| Strip test | | | |
| Strip | Test-pads are well fixed on foil | well fixed | passed |
| | Foil is clean | clean | passed |
| | Right sequence of test-pads | correct | passed |
| Material test | | | |
| Desiccant cap | Desiccant exists | exists | passed |
| | Cap closes tight | tight | passed |
| Tube and Label | No damages and contamination | intact | passed |
| | Correct and well legible print | correct | passed |
| Sleeve | Correct position and shrinkage | correct | passed |

This lot complies with our specification, in accordance with GMP.

Date of release: 2018-08-27 / Production Site Lichtenfels, Germany

This certificate of analysis has been automatically generated and is valid without signature.

Elena Schwietz
Acting Manager Quality Control