



Analyticon Biotechnologies GmbH
35104 Lichtenfels, Germany

Certificate of Analysis

Product: CombiScreen® mALB / CREA **Art.-No.:** 94025
Filling: 25 strips/tube

Lot number: 2304/1130
Manufacturing date: 2023-02-02
Expiry date: 2024-08-01
Storage: Store between +2 and +25 °C.

| Item | Specifications | Results | QC |
|---|---|------------|--------|
| Functional test | | | |
| <i>Crea - mAlb</i> | | | |
| Analyzer test | | | |
| Test with different concentrations/contents | Results correspond to the specified release criteria | correspond | passed |
| Visual test | | | |
| Test with different concentrations/contents | Colourations correspond to label, after the specified reading time (60 s) | correspond | passed |
| Strip test | | | |
| Strip | Test-pads are well fixed on foil | well fixed | passed |
| | Foil is clean | clean | passed |
| | Right sequence of test-pads | correct | passed |
| Material test | | | |
| Desiccant cap | Desiccant exists | exists | passed |
| | Cap closes tight | tight | passed |
| Tube, Label, Box and Insert | No damages and contamination | intact | passed |
| | Correct and well legible print | correct | passed |
| Sleeve | Correct position and shrinkage | correct | passed |

This lot complies with our specification.

Date of release: 2023-02-03 / Production Site Lichtenfels, Germany

This certificate of analysis has been automatically generated and is valid without signature.

Elena Schwietz
Manager Quality Control