

BIOLAB Zrt.

1141 Budapest, Öv u. 43.

Tel: **221 96 14** Fax: **364 20 06**

www.biolab.hu E-mail: biolab@biolab.hu

Certificate of analysis

Name:	KOVÁCS INDOLE REAGENT for diagnostics purpose
Code Number:	KIR60100
Lot:	KIR230823203
Manufacture Date:	2023/08
Expiry Date:	2024/08

Prepared Reagent:

Clarity:	Transparent
Deposit:	Free from any precipitate

Microbiological assay:

According to instruction for use

Quality control

Test strains	Growth
<i>Escherichia coli</i> ATCC 25922	Positive
<i>Salmonella typhimurium</i> ATCC 14028	Negative

This product has been tested by Quality Control Laboratory and conforms to the specification contained in the relevant catalogue or to the specification agreed with the customer. This product was manufactured by Biolab Inc.

This certificate was issued electronically and is valid without signature.