

**BIOLAB Zrt.**

1141 Budapest, Öv u. 43.

Tel: **221 96 14** Fax: **364 20 06**

www.biolab.hu E-mail: [biolab@biolab.hu](mailto:biolab@biolab.hu)

## Certificate of analysis

|                   |   |
|-------------------|---|
| <b>Name:</b>      | <b>KOVÁCS INDOLE REAGENT</b><br>for diagnostics purpose |
| Code Number:      | KIR60100  |
| Lot:              | KIR290722199  |
| Manufacture Date: | 2022/07   |
| Expiry Date:      | 2023/07   |

### Prepared Reagent:

|          |                           |
|----------|---------------------------|
| Clarity: | Transparent               |
| Deposit: | Free from any precipitate |

### Microbiological assay:

|                                  |
|----------------------------------|
| According to instruction for use |
|----------------------------------|

### Quality control

| Test strains                             | Growth   |
|--|----------|
| <i>Escherichia coli</i> ATCC 25922       | Positive |
| <i>Salmonella typhimurium</i> ATCC 14028 | Negative |

This product has been tested by Quality Control Laboratory and conforms to the specification contained in the relevant catalogue or to the specification agreed with the customer. This product was manufactured by Biolab Inc.

**This certificate was issued electronically and is valid without signature.**