

**BIOLAB Zrt.**

1141 Budapest, Öv u. 43.

Tel: **221 96 14** Fax: **364 20 06**

www.biolab.hu E-mail: [biolab@biolab.hu](mailto:biolab@biolab.hu)

## Certificate of analysis

<b>Name:</b>	<b>KOVÁCS INDOLE REAGENT</b> for diagnostics purpose
Code Number:	KIR60100
Lot:	KIR251023204
Manufacture Date:	2023/10
Expiry Date:	2024/10

### Prepared Reagent:

Clarity:	Transparent
Deposit:	Free from any precipitate

### Microbiological assay:

According to instruction for use
----------------------------------

### Quality control

Test strains	Growth
<i>Escherichia coli</i> ATCC 25922	Positive
<i>Salmonella typhimurium</i> ATCC 14028	Negative

This product has been tested by Quality Control Laboratory and conforms to the specification contained in the relevant catalogue or to the specification agreed with the customer. This product was manufactured by Biolab Inc.

**This certificate was issued electronically and is valid without signature.**