



# Certificate of Analysis

**Product:** CombiScreen® mALB / CREA **Art.-No.:** 94025  
**Filling:** 25 strips/tube

**Lot number:** 2695/1927  
**Manufacturing date:** 2024-06-27  
**Expiry date:** 2025-12-26  
**Storage:** Store between +2 and +25 °C.

Item	Specifications	Results	QC
<b>Functional test</b>			
<i>Crea - mAlb</i>			
<b>Analyzer test</b>			
Test with different concentrations/contents	Results correspond to the specified release criteria	correspond	passed
<b>Visual test</b>			
Test with different concentrations/contents	Colourations correspond to label, after the specified reading time (60 s)	correspond	passed
<b>Strip test</b>			
Strip	Test-pads are well fixed on foil	well fixed	passed
	Foil is clean	clean	passed
	Right sequence of test-pads	correct	passed
<b>Material test</b>			
Desiccant cap	Desiccant exists	exists	passed
	Cap closes tight	tight	passed
Tube, Label, Box and Insert	No damages and contamination	intact	passed
	Correct and well legible print	correct	passed
Sleeve	Correct position and shrinkage	correct	passed

This lot complies with our specification.

**Date of release:** 2024-07-15 / Production Site Lichtenfels, Germany

This certificate of analysis has been automatically generated and is valid without signature.

Elena Schwietz  
Manager Quality Control