

PRODOTTO / PRODUCT : EnteroPluri-Test

LOTTO / BATCH: 012325021

DATA PRODUZIONE/PRODUCTION DATE: 23.01.2025

DATA SCADENZA / EXP. DATE : 2025.09.20

CARATTERISTICHE BIOLOGICHE / BIOLOGICAL CHARACTERISTICS:

Tempo di incubazione / Incubation time	18-24 ore/ hours													
Temperatura / Temperature	36 °C ± 1 °C													
Modalità di incubazione / Procedure of incubation	$\sqrt{O_2}$ CO ₂													
Aggiunta di KOH 40% / Adding of KOH 40%	2 gocce/drops													
Aggiunta di α -naftolo / Adding of α -naphthol	3 gocce/drops													
Aggiunta di Kovac's / Adding of Kovac's	3-4 gocce/drops													

	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15
Microrganismi	GLU	GAS	LYS	ORN	H2S	IND	ADO	LAC	ARA	SOR	VP	DUL	PA	UR	CIT
<i>Escherichia coli</i> ATCC 25922															
Theoretical results	+	±	+	+	-	+	-	+	+	+	-	-	-	-	-
Results of conformity	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒
<i>Citrobacter freundii</i> ATCC 8090															
Theoretical results	+	+	-	±	+	-	-	±	+	+	-	±	-	-	±
Results of conformity	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒
<i>Proteus mirabilis</i> ATCC 25933															
Theoretical results	+	±	-	+	+	-	-	-	-	±	-	+	+	±	±
Results of conformity	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒
<i>Enterobacter cloacae</i> ATCC 13047															
Theoretical results	+	+	-	+	-	-	±	±	+	+	+	±	-	±	+
Results of conformity	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒
<i>Klebsiella pneumoniae</i> ATCC 13883															
Theoretical results	+	±	+	-	-	-	+	+	+	+	±	+	-	±	+
Results of conformity	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒
<i>Salmonella typhimurium</i> ATCC 14028															
Theoretical results	+	±	+	±	+	-	+	-	+	+	-	-	-	-	±
Results of conformity	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒

+ Reazione positiva / Positive reaction - Reazione negativa / Negative reaction ± Reazione variabile / Variable reaction

LOTTO / BATCH : Idoneo / Approved

Non Idoneo / Not approved

DATA / DATE : 06.02.2025

Responsabile Controllo Qualità
Quality Control Manager
(This document has been
established electronically and is valid without
signature)