

CERTIFICATE OF ANALYSIS

PRODUCT	PB5039A	COLUMBIA AGAR + SHEEP BLOOD PLUS
LOT NUMBER	6237427	
EXPIRY DATE	2025.05.15	
TEST DATE	2025.02.21	
REPORTING DATE	2025.02.24	

General Characteristics	Results	Specification
Colour	Conforms	Traffic red
Appearance	Conforms	Opaque
pH	7.4	7.1 -7.5
Packaging / Presentation	Conforms	Label & Print check
Cont. check @ 20-25 & 30-35°C for >= 72h	Conforms	Within Limits

Microbiological Performance	Control c.f.u	Test Result	Specification
Streptococcus pyogenes ATCC®12344	79	67	1-2mm, light grey cols & beta-haemolysis
Staphylococcus aureus ATCC®25923	<10000	Conforms	Good growth, white colonies
Staphylococcus aureus ATCC®6538	<10000	Conforms	Good growth, yellow cols & haemolysis
Streptococcus pneumoniae ATCC®6305	<10000	Conforms	Good growth, dark grey cols & alpha haem
Pseudomonas aeruginosa ATCC®9027	<10000	Conforms	Good growth, grey shiny cols
Escherichia coli ATCC®8739	<10000	Conforms	Good growth, dark grey cols

For positive strains, colony count is greater than or equal to 50% of the control medium.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were obtained at the time of release.

This certificate is produced electronically and valid without a signature

The quality control methods meet requirements of ISO 11133.