

Liofilchem S.r.l.	CERTIFICATO CONTROLLO QUALITÀ QUALITY CONTROL CERTIFICATE	N° 78618 - 78619 Revisione/ Revision 8 _ 14.09.2021 Pag. 1 di 1
------------------------------------	--	---

PRODOTTO / *PRODUCT* : **EnteroPluri-Test**

LOTTO / *BATCH*: **040925018**

DATA PRODUZIONE/*PRODUCTION DATE*: **09.04.2025**

DATA SCADENZA / *EXP. DATE* : **2025.12.05**

CARATTERISTICHE BIOLOGICHE / *BIOLOGICAL CHARACTERISTICS*:

Tempo di incubazione / <i>Incubation time</i>	18-24 ore/ <i>hours</i>	
Temperatura / <i>Temperature</i>	36 °C ± 1 °C	
Modalità di incubazione / <i>Procedure of incubation</i>	√ O ₂ CO ₂	
Aggiunta di KOH 40% / <i>Adding of KOH 40%</i>	2 gocce/drops	
Aggiunta di α-naftolo / <i>Adding of α-naphthol</i>	3 gocce/drops	
Aggiunta di Kovac's / <i>Adding of Kovac's</i>	3-4 gocce/drops	

	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15
<i>Microrganisms</i>	GLU	GAS	LYS	ORN	H2S	IND	ADO	LAC	ARA	SOR	VP	DUL	PA	UR	CIT
<i>Escherichia coli</i> ATCC 25922															
<i>Theoretical results</i>	+	±	+	+	-	+	-	+	+	+	-	-	-	-	-
<i>Results of conformity</i>	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒
<i>Citrobacter freundii</i> ATCC 8090															
<i>Theoretical results</i>	+	+	-	±	+	-	-	±	+	+	-	±	-	-	±
<i>Results of conformity</i>	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒
<i>Proteus mirabilis</i> ATCC 25933															
<i>Theoretical results</i>	+	±	-	+	+	-	-	-	-	-	±	-	+	+	±
<i>Results of conformity</i>	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒
<i>Enterobacter cloacae</i> ATCC 13047															
<i>Theoretical results</i>	+	+	-	+	-	-	±	±	+	+	+	±	-	±	+
<i>Results of conformity</i>	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒
<i>Klebsiella pneumoniae</i> ATCC 13883															
<i>Theoretical results</i>	+	±	+	-	-	-	+	+	+	+	±	+	-	±	+
<i>Results of conformity</i>	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒
<i>Salmonella typhimurium</i> ATCC 14028															
<i>Theoretical results</i>	+	±	+	±	+	-	-	-	+	+	-	-	-	-	±
<i>Results of conformity</i>	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒	☒

+ **Reazione positiva / *Positive reaction***

- **Reazione negativa / *Negative reaction***

± **Reazione variabile / *Variable reaction***

LOTTO / *BATCH* : ☒ Idoneo / *Approved*

☐ Non Idoneo / *Not approved*

DATA / *DATE* : 30.04.2025

Responsabile Controllo Qualità
Quality Control Manager
(This document has been
established electronically and is valid without
signature)