

Certificate of Analysis

Product:	RAPID-SCAN 6-SN +	Art.-No.: 01-04-11
		Filling: 100 strips/tube
Lot number:	2971/2724	
Manufacturing date:	2025-10-28	
Expiry date:	2027-10-27	
Storage:	Store between +2 and +30 °C.	

Item	Specifications	Results	QC
Functional test			
<i>Glu - Pro - Bld - Nit - pH - SG</i>			
Negative/normal test	negative/normal	negative/normal	passed
Test with different concentrations/contents	Colourations correspond to label, after the specified reading time (60 s)	correspond	passed
Strip test			
Strip	Test-pads are well fixed on foil Foil is clean Right sequence of test-pads	well fixed clean correct	passed passed passed
Material test			
Desiccant cap	Desiccant exists Cap closes tight	exists tight	passed passed
Tube and Label	No damages and contamination Correct and well legible print	intact correct	passed passed
Sleeve	Correct position and shrinkage	correct	passed

This lot complies with our specification.

Date of release: 2025-11-05 / Production Site Lichtenfels, Germany

This certificate of analysis has been automatically generated and is valid without signature.

Elena Schwietz
 Manager Quality Control