



Analyticon Biotechnologies GmbH
35104 Lichtenfels, Germany

Certificate of Analysis

| | | |
|----------------------------|------------------------------|---------------------------------|
| Product: | RAPID-SCAN 7SYS + | Art.-No.: 01-04-41 |
| | | Filling: 100 strips/tube |
| Lot number: | 3059/3046 | |
| Manufacturing date: | 2026-04-17 | |
| Expiry date: | 2028-04-16 | |
| Storage: | Store between +2 and +30 °C. | |

| Item | Specifications | Results | QC |
|---|--|-----------------|--------|
| Functional test | | | |
| <i>Ket - Glu - Pro - Bld - Nit - Leu - pH</i> | | | |
| Analyzer test | | | |
| Negative/normal test | negative/normal | negative/normal | passed |
| Test with different concentrations/contents | Results correspond to the specified release criteria | correspond | passed |
| Visual test | | | |
| Negative/normal test | negative/normal | negative/normal | passed |
| Test with different concentrations/contents | Colourations correspond to label, after the specified reading time (60 / 60-120 s) | correspond | passed |
| Strip test | | | |
| Strip | Test-pads are well fixed on foil | well fixed | passed |
| | Foil is clean | clean | passed |
| | Right sequence of test-pads | correct | passed |
| Material test | | | |
| Desiccant cap | Desiccant exists | exists | passed |
| | Cap closes tight | tight | passed |
| Tube and Label | No damages and contamination | intact | passed |
| | Correct and well legible print | correct | passed |
| Sleeve | Correct position and shrinkage | correct | passed |

This lot complies with our specification.

Date of release: 2026-04-24 / Production Site Lichtenfels, Germany

This certificate of analysis has been automatically generated and is valid without signature.

Elena Schwietz
Manager Quality Control