



Analyticon Biotechnologies GmbH  
35104 Lichtenfels, Germany

## Certificate of Analysis

**Product:** RAPID-SCAN 11SYS + **Art.-No.:** 01-04-31  
**Filling:** 100 strips/tube  
**Lot number:** 3053/2992  
**Manufacturing date:** 2026-03-31  
**Expiry date:** 2028-03-30  
**Storage:** Store between +2 and +30 °C.

| Item                                                                 | Specifications                                                                     | Results         | QC     |
|----------------------------------------------------------------------|------------------------------------------------------------------------------------|-----------------|--------|
| <b>Functional test</b>                                               |                                                                                    |                 |        |
| <i>Bil - Uro - Ket - Asc - Glu - Pro - Bld - pH - Nit - Leu - SG</i> |                                                                                    |                 |        |
| <b>Analyzer test</b>                                                 |                                                                                    |                 |        |
| Negative/normal test                                                 | negative/normal                                                                    | negative/normal | passed |
| Test with different concentrations/contents                          | Results correspond to the specified release criteria                               | correspond      | passed |
| <b>Visual test</b>                                                   |                                                                                    |                 |        |
| Negative/normal test                                                 | negative/normal                                                                    | negative/normal | passed |
| Test with different concentrations/contents                          | Colourations correspond to label, after the specified reading time (60 / 60-120 s) | correspond      | passed |
| <b>Strip test</b>                                                    |                                                                                    |                 |        |
| Strip                                                                | Test-pads are well fixed on foil                                                   | well fixed      | passed |
|                                                                      | Foil is clean                                                                      | clean           | passed |
|                                                                      | Right sequence of test-pads                                                        | correct         | passed |
| <b>Material test</b>                                                 |                                                                                    |                 |        |
| Desiccant cap                                                        | Desiccant exists                                                                   | exists          | passed |
|                                                                      | Cap closes tight                                                                   | tight           | passed |
| Tube and Label                                                       | No damages and contamination                                                       | intact          | passed |
|                                                                      | Correct and well legible print                                                     | correct         | passed |
| Sleeve                                                               | Correct position and shrinkage                                                     | correct         | passed |

This lot complies with our specification.

**Date of release:** 2026-04-08 / Production Site Lichtenfels, Germany

This certificate of analysis has been automatically generated and is valid without signature.

Elena Schwietz  
Manager Quality Control